



Product Safety Recall



Electrical Safety Recall

A power supply sold either as an accessory for, or bundled with, Liberator's 8 and 10 inch Rugged devices.

Only the power supply is under recall.

Purchased from Liberator or obtained through Speech Pathologists or NDIS plan managers.

Dates sold: 1 May 2020 to 31 March 2023

Identification:

Brand: **HUANAN**, Model No: **HNT-S520RZ**

Input: 100-240V~50/60Hz, 0.3A.

Output: 5.0V DC, 2.0A, 10.0 Watts.

Why the product is recalled: This product does not comply with the relevant electrical standard. There is no insulation on the base of the contact pins.

Hazard: Risk of serious injury or death from electric shock.

Action required:

Stop using this product immediately and contact Liberator to arrange a replacement power supply.

For further information

Contact [Liberator](https://www.liberator.net.au) on 02 9124 9945 or email info@liberator.net.au.

PRA 2023/19995

See [productsafety.gov.au](https://www.productsafety.gov.au) for
Australian product recall information!